



EUROPEAN SOUTHERN OBSERVATORY

Organisation Européenne pour des Recherches Astronomiques dans l'Hémisphère Austral
Europäische Organisation für astronomische Forschung in der südlichen Hemisphäre

ESO - European Southern Observatory
Karl-Schwarzschild Str. 2, D-85748 Garching bei München

Very Large Telescope

HAWK-I Calibration Plan

Doc. No. VLT-PLA-ESO-14800-3214

Issue 90, 18 Feb 2012

Prepared G. Carraro and the HAWK-I IOT
Name Date Signature

Approved C. Dumas
Name Date Signature

Released A. Kaufer
Name Date Signature

Change Record

Issue/Rev.	Date	Section/Parag. affected	Reason/Initiation/Documents/Remarks
1	03-02-04	All	1st version, prepared for PDR
2	26-10-04	All (only minor changes)	revision for FDR
3	23-03-07	All - monitoring removed	revision for PAE
81	02-12-07	All - only minor changes	Release for start of operations
82	08-06-08	Skyflats and zp sections	for P82
	22-06-08	Several typos	for P82
83	19-12-08	Several updates	for P83
84	27-06-09	Several updates	for P84
85	09-12-09	Updates	for P85
86	07-07-10	Minor updates	for P86 Phase II
88.2	11-06-11	Minor updates	for P88 Phase I and II
89	01-09-11	Minor updates	for P89 Phase I and II
90	18-02-12	—	for P90 Phase I and II

Contents

1	Introduction	1
1.1	Scope and content	1
1.2	Absence of internal calibration unit	2
1.3	References	2
1.4	Glossary	2
2	SCIENTIFIC DATA CALIBRATIONS	3
2.1	Summary: Science Data Calibrations	3
2.2	Darks	4
2.3	Twilight Flat-fields	5
2.4	Photometric calibrations	7
2.4.1	HAWK-I fundamental standards	7
2.4.2	HAWK-I 2MASS calibrator fields	8
2.4.3	Daily: Zero points and extinction coefficients monitor	8
2.4.4	Monthly: colour terms and extinction coefficients reference values	10
2.5	Illumination Frames	11
2.6	Astrometry	12
3	MAINTENANCE CALIBRATIONS	13
3.1	Summary: Technical Calibrations	13
3.2	Detector Characterization and Monitoring	14
3.3	Filter position monitoring	15
3.4	Flexure and Plate Scale	16
A	Daily Calibration Time Estimates	17

1 Introduction

1.1 Scope and content

This document describes the calibration procedures for the HAWK-I instrument at the VLT.

The document is intended for observatory staff and interested users. Users not interested in all the details of the operations can simply refer to the calibration summary in the HAWK-I User Manual.

The basic calibration of science data requires a set of calibrations obtained shortly before or after the observations and covering the specific instrument set-up used. The calibration and monitoring of the entire instrument, on the other hand, requires a large set of calibrations covering all offered set-ups and obtained periodically.

In practice, for instruments like HAWK-I which offer only few standard set-ups both tasks are combined. The science data are calibrated during night and day time, typically in a 24h window of the science observations. In addition, a standard set of calibrations is performed regularly (daily to weekly depending on their nature). Combined, these calibrations allow to monitor and calibrate the entire instrument. Finally, a few monitoring calibrations are obtained on a monthly basis to observe long range trends of the instrument.

For each calibration task this document defines the:

- **Responsible** group to carry out the calibrations,
- **Phase** when the calibrations have to be carried out (e.g. day or night time),
- **Frequency**, i.e., how often the calibration task has to be carried out,
- **Purpose** of the calibration,
- **Procedure** or the way how the calibration task is carried out,
- **Outputs**, i.e. the Pipeline data products, the Quality Control (QC) parameters, and/or the keywords entered into the VLT engineering data stream ("FITSLOG") that are being produced by the calibration task,
- **Prepared OBs** to carry out the task (and their location in a corresponding **OT queues**),
- **Prepared Templates** to carry out the task (if no prepared OB exists)
- **Duration**, i.e. the estimated time to execute the calibration task,
- **Pre-requisites**, i.e. the possible dependencies on instrumental or sky conditions, or other calibration tasks.

The appendix of the document provides an estimate of the expected typical daily calibration times for HAWK-I .

1.2 Absence of internal calibration unit

The experience with ISAAC and WFCAM (UKIRT) showed that the calibration of NIR imaging mode data does not necessarily require an internal calibration source. Accordingly, HAWK-I, a pure NIR imager, was not designed to include a calibration unit. None of the calibrations described in this document require internal lamp exposures. However, some calibrations use exposures of the uniform (thermal) illumination of the Nasmyth screen.

The issue of flat-fielding on sky is addressed in Sect. 2.3.

1.3 References

none

1.4 Glossary

Day/Night time Calibrations: Usually calibrations are done during day time following the night of observations. This implies a typical delay between science observations and day time calibrations of a few hours. Upon request, night time Calibration can be provided. Their execution time is charged against the time allocated to the science program, unless explicitly listed in the calibration plan.

Template is a set of instructions for the set-up of the instrument and detector followed by the performance of a specific standard operation. The templates are specially devised sequences used to simplify the execution of frequently performed instrument operations and calibrations.

Template Signature File: is a file describing for a given template its parameters, their default values and their allowed range of values.

Observation Block: it is a collection of templates to be executed as a single block in order to obtain a coherent set of data. It consists of the potential target information, a set of templates, parameter files (if needed) for the templates, a list of execution conditions/requirements, and, potentially, comments. Constructing Observation Blocks is part of the Phase II Proposal Preparation Process.

2 SCIENTIFIC DATA CALIBRATIONS

This section of the HAWK-I calibration plan summarizes the HAWK-I calibrations that have to be collected (and with which frequency) in order to allow to remove the instrumental signatures from the scientific data.

These calibrations are accessible as standard calibrations to the user.

2.1 Summary: Science Data Calibrations

HAWK-I – Calibration Plan			
TO BE UPDATED COMM1			
Calibration	number	frequency	comments / purpose
Darks	10 exp. / DIT-NDIT	daily	for each combination of DIT and maximum associated N
Darks	5 exp. / DIT-NDIT	daily	for each combination of DIT and maximum associated N
Twilight Flat-fields	1 set / filter	daily	broad-band filters (best effort basis)
	1 set / filter	as needed	for narrow-band filters
Zero points	1 set / (broad-band) filter	daily	zero points only when CLR or PHO
Zero points and extinction coefficients monitor	2 set / (broad-band) filter	daily	zero points and extinction coefficients when PHO
Colour terms	1 set	monthly	broad-band filters only (best effort basis)
Extinction coefficients reference values	1 set	monthly	broad-band filters only (best effort basis)
Illumination frames	1 set	monthly	zero point variations across the field
Astrometric calibration	1 + 1 sets	monthly	plate scale, distortions
Detector characteristics	1 set	monthly	RON, dark current, linearity, gain, fixed pattern noise, int

2.2 Darks

- **Responsible:** Science Operations
- **Phase:** Day time
- **Frequency:** Daily with several standard settings and as required for every DIT and associated maximum NDIT combination used during the previous night.
- **Purpose:** Calibration, read-out noise checks, hot pixel detection
- **Procedure:** Run the template, check results, report in night log. Darks are taken for each DIT and associated maximum NDIT . Ten Darks are taken for each DIT and associated maximum NDIT setting < 120 s. Five Darks are taken for each DIT and associated maximum NDIT setting >120 s. Windowed darks are provided in case FastPhot is used.
- **Outputs:** pipeline products:
 - Master dark frame
 - Hot pixel mapQC parameters:
 - Readout Noise: mean/median/rms
- **Prepared OBs:** calob based on template HAWKI_img_cal_Darks
- **OT queue:** N/A
- **Pipeline support:** Recipe hawki_img_dark can be used to produce a master dark frame. See pipeline manual for details.
- **Duration:** variable, experience showed it to be from 45 min to 2h.
- **Pre-requisites:** none
- **See also:**

2.3 Twilight Flat-fields

Flat-fielding is necessary to correct for several effects: to correct for pixel-to-pixel variations (each pixel of the array has a slightly different quantum efficiency than its neighbours); to correct for large scale variations of the quantum efficiency of the detectors; and in the case of HAWK-I to correct for the varying efficiency between the 128 read-out channels of the four arrays. The flat-field can also be used to compensate for vignetting of the optics, filter defects or dust on the optical elements. All above effects are multiplicative, and can ideally be removed by dividing a scientific exposure by the processed flat-field calibration frame.

Further additive effects exist (scattered light, fringing due to night sky emission lines, ...) but they do not change the signal of the astronomical object under study, i.e. do not need to be corrected for. They are, however, often difficult to disentangle from the multiplicative component and can thus limit the achievable correction.

Ideally, if the illumination would be homogeneous, flat-fielding would correct for all the multiplicative effects. For this purpose, the flat-field frames should be of a much higher S/N than the science images in order not to add noise to the latter.

Flat-fields for wide-field NIR imager are difficult to obtain given the restricted possibilities to homogeneously illuminate the field of view of the instrument. Wide-field imagers on 4m-class telescopes (WIRC/Palomar, OMEGA2000/Calar Alto, FLAMINGOS/KPNO, ISPI/CTIO) all use Dome flat-fields in addition of twilight flat-fields. However, ISAAC (with a ten times smaller field-of-view than HAWK-I), as well as WFCAM (UKIRT) only rely on twilight flat-fields in imaging mode and achieve $\sim 3\%$ flatness over their fields.

For HAWK-I, the possibility to illuminate the Nasmyth calibration screen has been considered. The location relatively close to the focal plane would require a very homogeneous illumination of the screen if it was aimed at correcting for the large scale variations. No affordable technical solutions exist to realize this. Lower constraints on the homogeneity of the illumination of the screen would allow to use 'internal' flat-fields to correct for small scale variations, but these are also well corrected by twilight flat-fields.

Thus, HAWK-I only relies on twilight flat-fields. These are sufficient to correct (up the $\sim 3\%$ level) for any pixel-to-pixel and small scale variations. The twilight flat-fields are also used to determine dead pixels and to adjust (harmonize) the gain between the four detectors of the mosaic. Zero point variations across the field are corrected by the illumination calibrations (see Sect. 2.5).

We did extensive tests and found that normal skyflats can be used in case FastPhot observations are done. There is no need for windowed flats. See the manuals for more details.

- **Responsible:** Science Operations
- **Phase:** Night time (morning and evening twilight)
- **Frequency:** Daily, depending on filter used. It is probably not strictly necessary to take flats every day, twilight flats might typically be used over several days.

- **Purpose:** to correct for pixel-to-pixel gain variations and small scale variations, to compute a bad pixel map and to harmonize the gain between the four detectors.
- **Procedure:** The acquisition of twilight flats with HAWK-I is similar to the NACO ones. It is usually possible to do one NB filters and a few BB filters in a single twilight. However, the window of opportunity for obtaining these flats lasts only a few minutes. Because star trails affect the flats we recommend to have 20 frames per filter, thus reducing the number of filters that can be done in a single twilight.

Skyflats for almost all filters are taking tracking at empty field. The only exception are red NB filters (see below.) The corresponding OBs are named SkyFlat-HHMM-BB for BB filters and SkyFlat-HHMM-NB for the NB filters and are located in the HAWKI TWILIGHT FLATS queue.

Red NB filters (BrG, H2 and NB2090) skyflats are taking without preset and tracking. The corresponding OB is named Flat-sequence-NB- red and are also located in the HAWKI TWILIGHT FLATS queue.

The user will then be prompted to set up the instrument. It is usually possible to start the twilight flats in NB 30 minutes before sunset, in Ks about 2 minutes after sunset, and in H, J, and Y about 12 minutes after sunset. In the evening, the procedure is to measure the number of counts on the RTD, and start when the counts reach ~ 25.000 ADU. The template is to be aborted when the number of counts is no longer decreasing. This will usually takes between 15 to 25 exposures.

In the morning, the procedure is the reverse one to the night – only H,J, and Y can be flat-fielded before the telescope is closed (10 min before sunrise).

- **Outputs:** pipeline products:
 - Master flat-field frames
 - Cold pixel mapQC parameters:
- **Prepared OBs:** see above, depending on the filter.
- **OT queue:** HAWK-I TWILIGHT FLATS
- **Pipeline support:** Recipe `hawki_img_flat` can be used to produce a flat field frame. See pipeline manual for the details.
- **Duration:** start ~ 1 h before sunset, ends ~ 20 min after sunset
- **Pre-requisites:** Clear / Photometric. Cloudy conditions do not allow to take proper twilight flats
- **See also:**

2.4 Photometric calibrations

2.4.1 HAWK-I fundamental standards

The following are intended as guidelines for the observations of the HAWK-I fundamental standard stars.

- At the beginning of the night one standard star is chosen to control the night quality. Basically a measure of the zero-point is obtained, and compared with previous nights. They have been chosen to have similar (blue) colours and thus the zero points measured with them are comparable. If the night is found to be photometric, a second standard at high airmass is taken to confirm the zeropoint and get an estimate of the extinction.
- The A0 star FS-5 is considered the HAWK-I fundamental standard, and it is observed whenever possible
- In poor seeing photometric nights stars at different airmasses are observed (instead of going idle)
- For proper determination of colour equations the orange and the red stars are observed (but not the reddened ones)
- The 2MASS fields, together with as many as possible of the UKIRT stars of different colours and at different airmasses, are observed for a proper determination of the transformation between 2MASS and HAWKI system.

Month	Star	RA	Dec	K	J-H	H-K	Spectra
	FS 1	00 33 54.48	-12 07 58.1	12.964 (0.003)	0.388 (0.003)	0.057 (0.004)	DK-G
	FS 3	01 04 21.63	+04 13 36.0	12.823 (0.003)	-0.111 (0.003)	-0.089 (0.005)	sdB
Dec-Jan	FS 5	01 54 34.65	-06 46 00.4	12.339 (0.005)	0.004 (0.002)	-0.005 (0.002)	A0
	FS 6	02 30 16.64	+05 15 51.1	13.382 (0.005)	-0.059 (0.005)	-0.069 (0.005)	DA3
Jan	FS 10	03 48 50.20	-00 58 31.2	14.983 (0.011)	-0.104 (0.006)	-0.118 (0.010)	DA2
	FS 114	04 19 41.72	+16 45 22.4	13.445 (0.006)	0.604 (0.002)	0.337 (0.006)	M7 V
Feb	FS 12	05 52 27.66	+15 53 14.3	13.916 (0.006)	-0.115 (0.003)	-0.094 (0.005)	sdO
Mar-Apr	FS 14	07 24 14.40	-00 33 04.1	14.198 (0.008)	-0.065 (0.007)	-0.031 (0.010)	O9-B2p
	FS 15	08 51 05.81	+11 43 46.9	12.348 (0.003)	0.337 (0.003)	0.054 (0.002)	G5 IV-V
	FS 16	08 51 15.01	+11 49 21.2	12.628 (0.004)	0.290 (0.003)	0.041 (0.002)	G1 V
	FS 17	08 51 19.31	+11 52 10.4	12.274 (0.002)	0.327 (0.003)	0.056 (0.002)	G4 V
May	FS 19	10 33 42.75	-11 41 38.3	13.782 (0.006)	-0.085 (0.005)	-0.115 (0.007)	DA2
	FS 128	11 05 10.40	+07 06 48.7	12.078 (0.004)	0.616 (0.003)	0.309 (0.005)	M5 V
	FS 20	11 07 59.93	-05 09 26.1	13.501 (0.010)	-0.038 (0.006)	-0.065 (0.008)	DA3
	FS 130	11 24 55.92	+34 44 38.5	12.252 (0.003)	0.596 (0.006)	0.099 (0.004)	K4
	FS 21	11 37 05.15	+29 47 58.4	13.147 (0.002)	-0.069 (0.004)	-0.090 (0.004)	DA3
Jun	FS 33	12 57 02.30	+22 01 52.8	14.254 (0.005)	-0.112 (0.002)	-0.103 (0.008)	DA1
Jul-Aug	FS 23	13 41 43.57	+28 29 49.5	12.375 (0.003)	0.527 (0.003)	0.066 (0.003)	G8 III
Aug-Sep	FS 142	18 29 51.26	+01 13 19.0	13.393 (0.011)	1.495 (0.005)	0.750 (0.006)	reddened?
	FS 143	18 29 53.79	+01 13 29.9	13.018 (0.010)	2.336 (0.011)	1.200 (0.005)	reddened?
	FS 145	18 30 05.81	+01 13 47.3	12.160 (0.006)	3.234 (0.049)	1.562 (0.009)	reddened?
Sep-Oct	FS 34	20 42 34.73	-20 04 34.8	13.000 (0.004)	-0.077 (0.003)	-0.074 (0.005)	DA3
Oct	FS 29	21 52 25.36	+02 23 20.7	13.311 (0.003)	-0.068 (0.002)	-0.070 (0.003)	DA3
	FS 31	23 12 21.60	+10 47 04.1	14.037 (0.007)	-0.130 (0.005)	-0.099 (0.005)	DA1
Nov	FS 32	23 16 12.37	-01 50 34.6	13.676 (0.007)	-0.107 (0.003)	-0.086 (0.005)	DAs

2.4.2 HAWK-I 2MASS calibrator fields

Besides the FS standards from the above list we are observing in parallel, in a best effort basis, a set of 2MASS fields so that we can cross-calibrate the HAWK-I photometry. These fields are adapted - in terms of field coverage and magnitude - from the 2MASS touchstone fields (<http://www.ipac.caltech.edu/2mass/releases/allsky/doc/sec3_{2d}.html>).

Field	RA	Dec	mag range	Comment
S90021	00 24 25.0	-02 00 10.6	10.56-13.54	
S90301	02 26 51.0	-39 32 33.6	10.78-11.54	
S90013	05 57 13.0	+00 19 08.0	10.01-12.01	
S90067	08 51 24.0	+11 52 24.8	9.76-12.66	include FS 16
S90217	12 01 59.0	-50 21 47.3	10.13-11.14	
S90868	15 00 29.0	-00 28 28.3	10.20-13.41	include T868-5385
S90547	18 51 18.0	-04 12 29.6	10.21-11.34	
S90813	20 41 20.0	-05 06 36.2	10.49-12.17	

2.4.3 Daily: Zero points and extinction coefficients monitor

Every night, the photometric zero points are determined.

- **Responsible:** Science Operations
- **Phase:** Night time
- **Frequency:** Nightly (visitor and service), 1 or 2 stars. Irrespective of the filters used during the night, the 4 broad band filters should be calibrated. **Zero points for the narrow band filters are NOT foreseen in the calibration plan.**
- **Purpose:** Zero point determination of the photometric calibration. Note that the checking for variations of the zero point across the field is not the purpose of this calibration but of the illumination frames (see Sect. 2.5). If the condition are CLR we observe only 1 standard field. In case the conditions are declared photometric, we provide a second standard field at high airmass to monitor the extinction coefficients, ideally soon after the OB which needs photometric conditions
- **Procedure:** The pre-defined OBs will position the photometric standard star at the center of the each detector. I.e. four images are obtained per filter.

The usual procedure (adopted by the calibration plan) is to measure the zero points at the beginning of the night, after the twilight flats, when the sky is still too bright for science observations, or at the end of the night. **Should the user require a good photometric calibration throughout the night, special calibration (charged against the user's time) need to be requested.**

Standards from the approved lists (UKIRT Faint standards or 2MASS touchstone fields) shall be used.

Always check the magnitude of the object. Stars with magnitudes fainter than 11.5 are preferred. If the counts in the central pixel are much higher than 25.000 ADU, then degrade the image quality (i.e. do a onecal, or badAO, instead of full active optics), or choose a fainter star.

- **Outputs:** pipeline products:
 - Reduced standard star image
 - QC parameters:
 - Zero points
- **Prepared OBs:** Many, based on the templates HAWKI_img_acq_Presetand HAWKI_img_cal_StandardStar
- **OT queue:** HAWK-I STANDARDS
- **Pipeline support:** Recipe hawki_img_zpoint
- **Duration:** Typically 30 min for 1 star in all 4 quadrants for all 4 filters.
- **Pre-requisites:** Photometric or clear conditions. (It might be useful under some clear/cloudy conditions (light uniform cirrus) to obtain zero points in order to get an idea of the cloud extinction.)
- **See also:** Illumination frames, Sect. 2.5

2.4.4 Monthly: colour terms and extinction coefficients reference values

Once per month, colour terms and extinction coefficients are monitored.

- **Responsible:** Science Operations
- **Phase:** Night time
- **Frequency:** Monthly (in service in bad seeing conditions). The 4 broad band filters should be calibrated.
- **Purpose:** Zero point, colour term and extinction coefficient determination of the photometric calibration.
- **Procedure:** The procedure is exactly the same as for the Daily zero point determination - except that the calibration is repeated three or four times throughout the night to follow the standard star from an airmass of ~ 1 down to an airmass of ~ 2 .
- **Outputs:** pipeline products:
 - Reduced standard star imageQC parameters:
 - Zero points
 - Colour terms
 - Extinction coefficientsThe latter two are not products of the pipeline, but of the QC trending.
- **Prepared OBs:** TBD, based on the templates `HAWKI_img_acq_Preset` and `HAWKI_img_cal_StandardStar`
- **OT queue:** `HAWK-I.Nighttime.Calibration`
- **Duration:** Three or four times 30 min through the night.
- **Pre-requisites:** Photometric conditions.
- **See also:** Illumination frames, Sect. [2.5](#)

2.5 Illumination Frames

The illumination frames are used to correct for large scale variations of the zero point across the field.

Only broad-band filters are supported in the calibration plan.

They complement the zero point calibration (see Sect. 2.4.3) and use the UKIRT faint standard star.

In the future the procedure might be replaced by observing 2MASS standard fields (avoiding multiple pointings). The photometric accuracy of 2MASS is typically 1.5% and sets the limit on the illumination correction.

- **Responsible:** Science Operations
- **Phase:** Night time
- **Frequency:** Once every month (frequency can be decreased if the stability is confirmed), or after intervention on the instrument.
- **Purpose:** modeled the zero point variations across the field
- **Procedure:** A standard field is typically dithered in a 5×5 pointing pattern with (automatically determined) ~ 400 pix offsets. The grid can be varied in the template. The variations in zero point across the field are measured and modeled.

The prepared OB will run through a set of pre-defined telescope offsets. Ensure that the number of counts and DIT/NDIT settings are such that the S/N is around a few hundreds on the star.

- **Outputs:** pipeline products:
 - Illumination correction mapQC parameters:
 - Illumination statistics (min, max, mean, rms)
- **Prepared OBs:** TBD, based on the template `HAWKI_img_tec_IlluFrame`
- **OT queue:** `HAWK-I.Nighttime.Calibration`
- **Duration:** The 5×5 grid in each quadrant was timed to need 1h40 per filter.
- **Pre-requisites:** Photometric, any seeing
- **See also:** Zero points, Sect. 2.4.3

2.6 Astrometry

The average distortion solution across the field is determined following the method outlined in Anderson et al. (2006, A&A 454, 1029) .

- **Responsible:** Science Operations
- **Phase:** Night time
- **Frequency:** Once every month (frequency can be decreased if the stability is confirmed), or after intervention on the instrument.
- **Purpose:** to determine the average distortion solution
- **Procedure:** to obtain a set of “dithered” images, observed in a regular grid of ~ 5 large (~ 200 pix) offsets. At each offset position, five exposures of 60s in (typically) Y or J are obtained.
- **Outputs:** pipeline products (supported in the future):
 - distortions
 - QC parameters:
- **Prepared OBs:** TBD based on the templates `HAWKI_img_acq_Preset` and `HAWKI_img_tec_Astrometry`
- **OT queue:** `HAWK-I.Nighttime.Calibration`
- **Duration:** 45–60 min
- **Pre-requisites:**
- **See also:** Flexure and Plate Scale, Sect. 3.4.

3 MAINTENANCE CALIBRATIONS

3.1 Summary: Technical Calibrations

This section of the HAWK-I calibration plan summarizes the HAWK-I *technical* calibrations that are being collected to complement the science data calibrations. They are used by ESO to compute instrument characteristics that are then provided to the user (e.g. field distortions, zero point variations, ...). These calibrations are performed less frequently than the science data calibrations. They are normally NOT accessible as special calibrations to the user.

Monitoring data are normally made available as QC parameters through QC web pages and FIT-SLOG keywords through the Paranal Autrep database with standard plots created on a daily basis.

3.2 Detector Characterization and Monitoring

The characteristics (read-out noise, dark current, bad pixels, linearity, persistence, ...) of the detector mosaic are monitored periodically.

This is partly achieved with the calibrations listed above (in particular with Darks, see Sect. 2.2). Yet, several characteristics (linearity, conversion factor, ...) require a set of internal flat-fields with various illumination in order to be determined. This is the purpose of this calibration. Note, however, that HAWK-I has no internal calibration source, i.e. the images are acquired on the dark Nasmyth screen in the Ks band.

- **Responsible:** Science Operations
- **Phase:** Daytime
- **Frequency:** Monthly
- **Purpose:** Characterize the Detector via the following parameters:
 - RON as a function of read-out mode
 - Dark current
 - Conversion factor
 - Bad pixel/columns map
 - Linearity
 - Gain
 - Fixed Pattern Noise
 - Inter-pixel capacitance correction
- **Procedure:** Take a series of flats, darks for the standard detector modes
- **Outputs:** pipeline products and associated template:
 - QC parameters:
 - RON (HAWKI_img_cal_Darks)
 - Dark current (HAWKI_img_cal_Darks)
 - Bad pixel/columns map (HAWKI_img_cal_TwFlats)
 - Linearity (HAWKI_img_tec_DetTest)
 - Gain (HAWKI_img_tec_DetTest)
 - Fixed pattern noise (HAWKI_img_tec_DetTest)
 - Inter-pixel capacitance correction (HAWKI_img_tec_DetTest)
- **Prepared OBs:** TBD, based on HAWKI_img_cal_Darks, HAWKI_img_cal_TwFlats, and HAWKI_img_tec_DetTest
- **OT queue:** HAWK-I.Daytime.Calibration
- **Duration:** $\sim 2h$
- **Pre-requisites:** Detectors have stabilized, i.e. been online for > 6 h.
- **See also:** Darks and Flats, Sects. 2.2, 2.3

3.3 Filter position monitoring

This calibration is used to exercise the filter wheel. The data taken are then used to monitor some of the internal “flat fields”. Since HAWK-I does not have a calibration’s unit, the “internal flats” correspond to long integrations on the Nasmyth shutter. They cannot be used to calibrate science data (they are not flat) but are useful to monitor the instrument’s transparency (throughput).

- **Responsible:** Science Operations
- **Phase:** Daytime
- **Frequency:** Monthly
- **Purpose:** to exercise the filter wheel and to determine the stability of the “internal flats”
- **Procedure:** One can give a list of filters for both filter wheels: FW1 and FW2. The template starts working through the list given for FW1. It sets a filter (e.g. J) and takes an image then move back to the filter defined as first in that list and moves back (i.e to J) and takes another image. It was originally meant to test the filter wheel reproducibility but not having e.g. a pinhole array, it is not possible. Instead we use this to take “internal flats” to monitor the transparency of the optics. Once the OB ran through the list of FW1, it then repeats the same procedure for FW2.
- **Outputs:** Median flux for a given filter (over the full field and some pre-defined sub-windows).
- **Prepared OBs:** TBD, based on `HAWKI_FW_test_daily`
- **OT queue:** `HAWK-I.Daytime.Calibration`
- **Duration:** ~ 30 min
- **Pre-requisites:** At most lower lights in the dome
- **See also:** Twilight Flat-fields

3.4 Flexure and Plate Scale

This calibration complements the first step of the astrometric calibration (see Sect. 2.6).

The absolute scale and flexures are determined. Note that the flexures are currently measured to be negligible (< 1 pix over 180 deg rotation; while typical rotations over one hour are < 30 degrees when not crossing meridian close to zenith).

- **Responsible:** Science Operations
- **Phase:** Night time
- **Frequency:** Monthly
- **Purpose:** to measure the instrument flexures and determine the absolute plate scale
- **Procedure:** point towards a standard astrometric field, configure the instrument, and acquire one image per position angle of the telescope adapter/rotator, while the later scans over 400 degrees in steps of 30 degrees (TBD).

Use standard astrometric fields from the approved lists.

- **Outputs:** pipeline products:
 - center of rotation
 - flexure measure
 - plate scale
 - rotator offset
 - WCSQC parameters:
- **Prepared OBs:** TBD based on the templates `HAWKI_img_acq_Preset` and `HAWKI_img_tec_Flexure`
- **OT queue:** `HAWK-I.Nighttime.Calibration`
- **Duration:** ~ 30 min
- **Pre-requisites:** science calibrations and average distortion map
- **See also:** Astrometry, Sect. 2.6

A Daily Calibration Time Estimates

Typical calibration times will first be recorded during commissioning. More detailed estimates for this section are expected from experience collected during the dry run and operations.

Overall, the time for daytime calibration is expected to be short. An estimate for daytime calibrations is:

Type	Set-ups	Estimated time per set-up	Total time
Darks	each DIT/NDIT combination	<1000s	45 min to 2h
Filter Wheel check	all filters		30 min

An estimate for night time calibrations is:

Type	Set-ups	Estimated total time
Twilight Flat-fields	all filters/twilight	80 min (from 1h before, until 20 min after sunset)
Zero points	1 star / quad / filter	30 min before astronomical twilight
Zero points +	as above at 3-4 airmasses	30 min per airmass; 2h in a night
Illumination frames	1 set / filter	100 min per filter
Astrometric calibration	1 set of 25 images	45-60 min

---oOo---